

.

---

**From:** Clemmer, Jill (DPH)  
**Sent:** Friday, October 15, 2010 3:30 PM  
**To:** Nassif, Julianne (DPH)  
**Cc:** 'Salemi, Charles (DPH)'; 'Lawler, Michael (DPH)'; Gu, Jane X. (DPH)  
**Subject:** Drug Lab - GHB Assistance

FYI – I've been asked to help Mike Lawler with HPLC analyses of suspect GHB samples – Drug Lab samples. Seems their HPLC system (the Agilent system we got from Tb back when) is kaput – the description of the problem Mike gave me is that the new repair parts have upgraded firmware that can't communicate with their software. I imagine that Agilent might upgrade their software for free, but I'm not aware of any action to pursue this by anyone in the Drug Lab. I may be forgetting some other issue with their LC – not sure...

I am willing to assist Michael with their GHB sample backlog using our Beckman HPLC system. Currently, Mike and I agreed to chip-away at the backlog weekly (one day per week) starting as soon as Mike has samples ready to go. Jane or I can setup the Beckman LC system; Mike would prep all of the samples, stds, mobile phases and be named on the reports.

This fly with you?

Jill

*Jill Clemmer  
Laboratory Supervisor  
Environmental Chemistry Laboratory  
Massachusetts DPH  
William A. Hinton State Laboratory Institute  
305 South St., room 309  
Jamaica Plain, MA 02130  
617-983-6657  
617-983-6662 FAX  
website: [www.mass.gov/dph](http://www.mass.gov/dph)*